



Infant Male & Female Catheterization Trainer

LF01035U

INSTRUCTION MANUAL





Figure 1



Figure 2

ABOUT THE SIMULATOR

The **Life/form**® Infant Male and Female Catheterization Trainer is an inexpensive 2-in-1 trainer that provides realistic practice of this difficult procedure. Actual product may vary slightly from photo. Nasco reserves the right to change product color, materials, or function as needed.

- Practice catheter insertion and position.
- Feel resistance and pressure as with a real patient
- Urine flows when catheter is in proper position
- Practice both male and female infant catheterization training
- Anatomically realistic with both male and female genitals
- Genitals made of soft, lifelike material
- Replaceable male and female genitals
- Lower half of infant only

LIST OF COMPONENTS

- Infant torso
- Female genitals set
- Male genitals set
- Infant intermittent catheter (8 FR, 6")
- 4-oz. squeeze bottle (125 ml)
- 2-oz. bottle of **Life/form**® lubricant
- Storage box
- Instruction manual

GENERAL INSTRUCTIONS FOR USE

Changing of the Male and Female Genitals

1. To install genitals, insert and slide bladder tube with genitals through opening in torso. Guide the bladder tube through the top cover and position genitals firmly in place within the torso. (See *Figure 1.*)
2. To remove genitals, make sure bladder tube cap is closed, then pull genitals through torso while gently pushing the bladder tube through the top cover. (See *Figure 1.*)

Inserting the Urinary Catheter

1. The trainer is designed for use with a standard 8 French latex or silicone intermittent catheter only. The silicone catheter may be easier to use as it is somewhat stiffer. Do NOT use a Foley catheter.
2. To fill the bladder tube, remove the luer cap and carefully inject desired amount of water using the squeeze bottle provided. Recap the bladder tube tightly until ready to begin actual training. (See *Figure 2.*)
3. Prop the simulator in a semi-reclining position with a diaper or paper towel underneath to catch leakage that may occur.
4. Lubricate the catheter with water-based lubricant, taking care not to block the drain holes.
5. Loosen the luer cap when ready to begin; **water will not flow when tube is tightly capped.**



Figure 3

6. Gently and slowly insert the catheter. Squeezing and releasing the tubing with your thumb and finger as you do so may help to start the flow. Stop insertion as soon as flow begins, and leave the catheter in place until it stops completely. (See Figure 3.)
7. Always remove the catheter immediately after training. **Leaving the catheter in place for extended periods, especially if it is the latex type, may permanently damage both the catheter and the simulator.**
8. When storing the trainer, allow the bladder tube to air dry by removing the luer cap. **Note:** The soft material of the trainer's catheter valve may sometimes fail to return to its fully closed position after catheterization, resulting in some leakage if the bladder tube is refilled immediately. If so, you may wish to begin catheter insertion before opening the luer cap. If leakage persists, remove the bladder tube cap and insert the catheter into the tube from the top. Gently tap the tip against the valve several times. Continued leaking may require replacement of the bladder tube.

CARE OF THE SIMULATOR

1. Normal surface soil can be removed with a mild soap and lukewarm water. Use Nasco Cleaner (LF09919U) to remove stubborn stains. Simply apply the cleaner to the soiled area and wipe with a clean, soft paper towel. **DO NOT** clean the simulator with solvents or corrosive materials, as they will cause damage.
2. **NEVER** place the simulator on any kind of printed paper or plastic. These materials, as well as ballpoint pens, will transfer indelible stains.

REPLACEMENT PARTS


LF01035(A)U	Replacement Genital Insert — Female
LF01035(B)U	Replacement Genital Insert — Male
LF01074U	Infant Intermittent Catheter (8 FR, 6")
LF01074(A)U	10 Infant Intermittent Catheters (8 FR, 6")
LF00985U	Life/form ® Lubricant Kit, 2-oz. Bottles (Pkg. of 6)
LF01221U	Replacement Bladder Tube
LF09919U	Nasco Cleaner

Actual product may vary slightly from photo. Nasco Healthcare reserves the right to change product color, materials, supplies, or function as needed.

Other Available *Life/form* Simulators

LF00698U	Adult Injectable Arm (White)	LF01038U	Episiotomy Suturing Set
LF00855U	Male Catheterization	LF01042U	Suture Kit
LF00856U	Female Catheterization	LF01062U	Pelvic, Normal & Abnormal
LF00901U	Prostate Examination	LF01069U	Cervical Effacement
LF00906U	Ostomy Care	LF01070U	Birth Station
LF00957U	Enema Administration	LF01084U	Sigmoidoscopic Examination
LF00961U	Intramuscular Injection	LF01095U	Blood Pressure Arm
LF00984U	Breast Examination	LF01142U	Auscultation Trainer
LF01005U	First Aid Arm	LF01143U	Testicular Exam
LF01008U	Intradermal Injection Arm	LF01150U	Male Catheter Model
LF01025U	Male Cath-Ed 1 Simulator	LF01151U	Female Catheter Model
LF01026U	Female Cath-Ed 2 Simulator	LF01152U	Male & Female Catheter
LF01027U	Peritoneal Dialysis	LF01193U	Special Needs Infant
LF01036U	Spinal Injection	LF04000U	GERI™/KERI™
LF01037U	Hemodialysis Practice Arm		Manikin Series

LF00855U



LF01193U



LF00856U



ONLINE:
EMAIL:
PHONE:
ADDRESS:

NascoHealthcare.com
lifeform@enasco.com
1.800.431.4310
901 Janesville Ave.
Fort Atkinson, WI 53538